


# Layout History

- Layout 2016-01, valid for 2016 data before May 27
- Layout 2016-02, valid for 2016 data since May 27
- scintillator working point at VThreshold = 30mV

Scintillator working point at  $V_{\text{Threshold}} = 30\text{mV}$

	HV (V)	Rate (kHz)
scint. 1	1650	
scint. 2	1500	
scint. 3	1550	
scint. 4	1500	
scint. 15	1750	
scint. 6	1650	
scint. 7	1900	
scint. 8	1600	

# Layout 2016-01




	VMON	IMON	V0
CH00	1	2	1650
CH01	1	0	1500
CH02	1	0	1550
CH03	2	0	1500
CH04	1	0	1750
CH05	1	0	1600
CH06	1	0	1650
CH07	1	0	1900

view from PMT side

Lab  
wall

**valid for 2016 data before May 27**

# Layout 2016-02



	VMON	IMON	V0
CH00	1	1	1650
CH01	1	0	1600
CH02	1	0	1550
CH03	1	0	1500
CH04	1	0	1750
CH05	1	0	1500
CH06	1	0	1650
CH07	1	0	1900

view from PMT side

valid for 2016 data since May 27

Lab  
wall